



## CERTIFICATE OF CONFORMITY

### CE EMC EN 55032 + EN 55024 Test Report

For the following information

Ref. File No.: C1M2011084  
(C1M1605220)

Product	Medical Grade Mobile Tablet PC
Model	MMS1200
Brand	MACTRON GROUP
Applicant	MACTRON GROUP CO., LTD.
Manufacturer	MACTRON GROUP CO., LTD.
Test Report Number	EM-E200669
Standards	EN 55032:2012 +AC:2013 AS/NZS CISPR 32:2013 EN 61000-3-2:2014 and EN 61000-3-3:2013 EN 55024:2010+A1:2015 (IEC 61000-4-2:2008, IEC 61000-4-3:2010, IEC 61000-4-4:2012, IEC 61000-4-5:2014, IEC 61000-4-6:2013, IEC 61000-4-8:2009, IEC 61000-4-11:2004)

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the council EMC directive 2014/30/EU.

The test data and results are issued on the EMC test report no. EM-E200669.

Signature

  
Ben Cheng/Manager

Date: 2020. 11. 10

Test Laboratory:

AUDIX Technology Corporation, EMC Department

NVLAP Lab. Code: 200077-0

TAF Accreditation No.: 1724

Web Site: [www.audixtech.com](http://www.audixtech.com)

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.