



## CERTIFICATE OF CONFORMITY

### ETSI EN 301 489-1/-3/-17/-19 Test Report

For the following information

Ref. File No.: C1M2011083  
(C1M1801003)

Product	Medical Grade Mobile Tablet PC
Model	MMS0700
Brand Name	MACTRON GROUP
Applicant	MACTRON GROUP CO., LTD.
Test Report Number	EM-E200665A
Standards	ETSI EN 301 489-1 V2.1.1: 2017-02 Final Draft ETSI EN 301 489-3 V2.1.1: 2017-03 ETSI EN 301 489-17 V3.1.1: 2017-02 Draft ETSI EN 301 489-19 V2.1.0: 2017-03 EN 55032:2015, Class B EN 61000-3-2:2014 and EN 61000-3-3:2013 EN 61000-4-3:2006 +A1:2008 +A2:2010, EN 61000-4-2:2009, EN 61000-4-4:2012, EN 61000-4-6:2014, EN 61000-4-11:2004+A1:2017, EN 61000-4-5:2014

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the Radio Equipment Directive 2014/53/EU. The test data & results are issued on the test report no. EM-E200665A.

Signature

  
Ben Cheng/Manager

Date: 2020. 11. 10

Test Laboratory:

AUDIX Technology Corporation, EMC Department

NVLAP Lab. Code: 200077-0

TAF Accreditation No.: 1724

Web Site: [www.audixtech.com](http://www.audixtech.com)

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.